

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK



0147

Accredited to
ISO/IEC 17025:2005

Agilent Technologies UK Limited

Issue No: 036

Issue date: 3 October 2008

Service Solutions Unit
610 Wharfedale Road
Winnersh Triangle
Wokingham
Berkshire
RG41 5TP
United Kingdom

Contact: Test & Measurement Contact Centre
Tel: +44 (0)7004 666666
Fax: +44 (0)7004 444555
E-Mail: contactcenter_UK@agilent.com
Website: www.agilent.co.uk

Calibration performed by the Organisations at the locations specified below

The locations covered by the organisation and their relevant activities are shown on Page 2 of this Schedule of Accreditation. Links to the relevant parts of the schedule for each type of calibration performed are included below.

Measured Quantity Instrument or Gauge	United Kingdom	Site Calibration	Germany
AC Current	▶	▶	
AC Resistance	▶		
AC Voltage	▶	▶	▶
Amplitude Modulation	▶	▶	
Capacitance	▶		
CISPR16 Pulse Response	▶		
Complex VRC / VTC (<i>iPIMMS</i>)	▶		
DC Current	▶	▶	▶
DC Resistance	▶	▶	▶
DC Voltage	▶	▶	▶
Directivity	▶		
Dissipation Factor	▶		
Distortion	▶		
Frequency	▶	▶	▶
Frequency Modulation	▶	▶	
Impedance (LISN calibration)	▶		
Pulse Characteristics	▶		
RF Attenuation	▶		
RF Calibration Factor	▶	▶	▶
RF Power	▶	▶	
RF Voltage	▶	▶	▶
Time Interval	▶	▶	
VRC/VSWR	▶	▶	▶






0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Laboratory locations:

Location details		Activity	Location code
Address Service Solutions Unit 610 Wharfedale Road Winnersh Triangle Wokingham Berkshire RG41 5TP	Local contact Test & Measurement Contact Centre. Tel: +44 (0) 7004 666666 Fax: +44 (0) 7004 444555 E-mail: contactcenter_UK@agilent.com	<u>Calibration</u> DC and LF electrical quantities RF and Microwave electrical quantities	 United Kingdom
			 Germany
Agilent Technologies Sales & Services GmbH & Co. KG Servicezentrum Herrenberger Straße 130 71034 Böblingen Germany	Local contact Mr Wolfgang Trester Tel: +49 (0) 7031 464 6569 Fax: +49 (0) 7031 464 6565 E-mail: wolfgang_trester@agilent.com	<u>Calibration</u> DC and LF electrical quantities RF and Microwave electrical quantities	 Germany
			 United Kingdom

Activities performed away from the locations listed above:

Location details		Activity	Location code
Customers' sites or premises Calibrations may be performed in an air-conditioned vehicle taken to the customers' sites or in suitable areas within the customers' premises. The customers' premises must be appropriate for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Local contact Test & Measurement Contact Centre. Tel: +44 (0) 7004 666666 Fax: +44 (0) 7004 444555 E-mail: contactcenter_UK@agilent.com	<u>Calibration</u> DC and LF electrical quantities RF and Microwave electrical quantities	Site calibration



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
DC VOLTAGE				United Kingdom
Specific Values	1 V to 1.1 V 10 V	1.2 μ V 0.8 ppm	Standard cells and solid-state reference standards can be measured on a fully automated system.	
Other Values	10 mV	23 ppm	Generation Measurement	
	0.1 V	2.5 ppm		
	1 V	0.8 ppm		
	10 V	0.8 ppm		
	100 V	0.8 ppm		
	1 kV	1.0 ppm		
	0 V *	0.1 μ V		
	0 to 10 V *	0.16 μ V		
	* With respect to a reference established using a low thermal emf short circuit			
	Up to 0.21 V	1.8 ppm + 0.3 μ V		
	0.21 V to 2.1 V	1.4 ppm + 0.3 μ V		
	2.1 V to 100 V	3.7 ppm		
	100 V to 1 kV	5.0 ppm		
DC CURRENT	0 A	1 pA	Appropriate to calibration of precision digital multimeters (open circuit input).	
	100 nA to 1 μ A	27 pA	Measurement only	
	1 μ A to 10 μ A	27 ppm to 13 ppm		
	10 μ A to 100 μ A	7 ppm to 4 ppm		
	100 μ A to 1 mA	7 ppm to 4 ppm		
	1 mA to 10 mA	7 ppm to 4 ppm		
	10 mA to 100 mA	7 ppm to 4 ppm		
	100 mA to 1 A	34 ppm to 15 ppm		
	1 A to 10 A	300 ppm to 47 ppm		
	10 A to 100 A	370 ppm to 570 ppm		
	100 A to 200 A	570 ppm		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)		Remarks	Location Code					
AC VOLTAGE	1 mV 10 mV 100 mV	10 Hz to 100 kHz	100 kHz to 1 MHz	Ratio step-down method	United Kingdom					
Voltage	Best Measurement Capability in ppm for specific values of AC voltages at the frequencies shown									
	10 Hz	40 Hz	1 kHz	10 kHz		30 kHz	50 kHz	100 kHz	500 kHz	1 MHz
100 mV		50	50	56		50	50	50	200	1000
300 mV		30	30			35	35	35	160	400
1 V	40	30	30	37		30	40	40	150	440
2 V		31	31	37			49	96	440	
3 V	55	40	40			40	40	40	160	390
10 V	50	30	30	120		30	35	35	160	390
20 V									120	360
30 V	45	30	30			30	35	35		
100 V	50	40	35	130		35	35	40		
	10 Hz	45 Hz	1 kHz	10 kHz		20 kHz	33 kHz	50 kHz	100 kHz	
300 V		50	50	50		50	60	60	80	
500 V									200	
700 V							150			
1000 V		45	45	45	45	65				



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code		
AC VOLTAGE (cont'd)				United Kingdom		
	Best Measurement Capability in % for other values of AC voltages over the frequency ranges shown					
Voltage range	10 Hz to 100 Hz	100 Hz to 30 kHz	30 kHz to 200 kHz		200 kHz to 500 kHz	500 kHz to 1 MHz
1 mV to 3.3 mV	0.15	0.13	0.19		0.35	0.70
3.3 mV to 10 mV	0.048	0.030	0.069		0.20	0.47
10 mV to 33 mV	0.038	0.023	0.050		0.15	0.36
33 mV to 100 mV	0.029	0.014	0.027		0.080	0.21
	Best Measurement Capability in ppm for other values of AC voltages over the frequency ranges shown					
	10 Hz to 40 Hz	40 Hz to 30 kHz	30 kHz to 200 kHz		200 kHz to 500 kHz	500 kHz to 1 MHz
0.1 V to 0.33 V	50	50	90		370	960
0.3 V to 1.1 V	50	50	90		360	960
1 V to 3.3 V	60	50	90		360	940
3 V to 11 V	60	45	90		360	940
10 V to 33 V 20 V max. ≥ 500 kHz	60	50	90		350	860
30 V to 110 V 100 kHz max. ≥ 100 V	60	50	90			
	10 Hz to 40 Hz	40 Hz to 20 kHz	20 kHz to 100 kHz			
100 V to 330 V	65	65	140			
300 V to 1.1 kV 700 V max. ≥ 50 kHz	65	80	220			
AC CURRENT						
Specific Values	<i>From 10 Hz to 5 kHz:</i> 10 µA 100 µA 1 mA 10 mA 100 mA <i>From 20 Hz to 10 kHz:</i> 10 mA 100 mA 1 A 10 A	0.087% 0.033% 0.033% 0.033% 0.033% 0.011% 0.011% 0.015% 0.027%	Voltage/resistance method Thermal transfer method			
Other Values	10 µA to 110 µA 20 Hz to 1 kHz 100 µA to 110 mA 20 Hz to 100 Hz 100 Hz to 5 kHz	0.4% to 0.1% 0.4% to 0.1% 0.23% to 0.05%				




0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location
				 United Kingdom
AC CURRENT (cont'd) Other values	100 mA to 1.05 A 20 Hz to 100 Hz 100 Hz to 5 kHz	0.40% to 0.10% 0.30% to 0.12%		United Kingdom
	1.05 A to 11 A 40 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.23% to 0.12% 0.32% to 0.17% 0.89% to 0.34%		
AC RESISTANCE				
Specific Values	1 Ω 5 Hz to 100 kHz 100 kHz to 1 MHz	0.17% 0.23%	Appropriate to the calibration of meters and standards with 4-terminal pair BNC connectors. Other configurations can be calibrated but the uncertainties may be increased.	
	10 Ω 5 Hz to 1 MHz 1 MHz to 13 MHz	0.14% 2.1%		
	100 Ω 5 Hz to 1 MHz 100 kHz, 500 kHz and 1 MHz 3 MHz 5 MHz 10 MHz 1 MHz to 10 MHz 10 MHz to <13 MHz At 13 MHz	0.16% 0.14% 0.33% 0.33% 0.33% 0.38% 0.41% 0.37%		
	1 k Ω 5 Hz to 1 MHz 100 kHz, 500 kHz and 1 MHz 3 MHz 5 MHz 10 MHz 13 MHz 1 MHz to 13 MHz	0.24% 0.21% 0.32% 0.32% 0.33% 0.34% 0.38%		
	10 k Ω 5 Hz to 1 MHz 100 kHz 500 kHz 1 MHz	0.24% 0.21% 0.21% 0.21%		
	100 k Ω 5 Hz to 500 kHz 100 kHz 500 kHz 1 MHz 500 kHz to 1 MHz	0.35% 0.33% 0.33% 0.39% 0.41%		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC RESISTANCE (cont'd)				United Kingdom
Other Values	1 Ω to 100 M Ω 20 Hz to 1 MHz	0.11%	The uncertainty varies in a complex manner. This is the lowest possible uncertainty within the stated range. The maximum frequency for resistance measurement above 100 k Ω is limited to f_{max} (kHz) = $10^5/R$ (k Ω).	
IMPEDANCE				
For calibration of Line Impedance Stabilisation Networks.	1 Ω to 200 Ω 20 Hz to 30 MHz	5%		
CAPACITANCE				
Specific Values	1 pF 5 Hz to 1 kHz 1 kHz to 1 MHz 1 MHz to 13 MHz	0.04% 0.38% 1.3%	For the calibration of meters and standards with 4-terminal pair BNC connectors. Other configurations can be calibrated but the uncertainties may be increased.	
	10 pF 5 Hz to 1 kHz 1 kHz 1 kHz to 1 MHz 1 MHz to 13 MHz	0.03% 0.026% 0.07% 0.21%		
	100 pF 5 Hz to 1 kHz 1 kHz 1 kHz to 1 MHz 1 MHz to 13 MHz	0.23% 0.024% 0.05% 0.19%		
	1000 pF 5 Hz to 1 kHz 1 kHz 1 kHz to 1 MHz 3 MHz 5 MHz 10 MHz 13 MHz 1 MHz to 13 MHz	0.25% 0.023% 0.04% 0.51% 0.51% 0.52% 0.53% 0.60%		
	10 nF 120 Hz 1 kHz 10 kHz 100 kHz	0.04% 0.035% 0.06% 0.16%		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
CAPACITANCE (cont'd)				United Kingdom
Specific Values	100 nF 120 Hz 1 kHz 10 kHz 100 kHz	0.04% 0.03% 0.10% 0.29%		
	1 μF 120 Hz 1 kHz 10 kHz	0.06% 0.03% 0.08%		
Other Values <i>Measurement only</i>	1 pF to 1 mF ($D < 0.1$) 20 Hz to 1 MHz	0.15%	The uncertainty varies in a complex manner. This is the lowest possible uncertainty in the stated range. The maximum frequency for capacitance measurement above 100 nF is limited according to $f_{max} \text{ (kHz)} = 10^{-4}/C \text{ (F)}$.	
DISSIPATION FACTOR	D (tan δ) values between 0 and 0.01 at the following nominal values of capacitance:		For the calibration of meters and standards with 4-terminal pair BNC connectors. Other configurations can be calibrated but the uncertainties may be increased.	
	1 pF 5 Hz to 1 MHz 1 MHz to 13 MHz	0.0004 0.0012		
	10 pF 5 Hz to 1 MHz 1 MHz to 13 MHz	0.0002 0.0012		
	100 pF 5 Hz to 1 kHz 1 kHz to 1 MHz 1 MHz to 13 MHz	0.0027 0.0004 0.0012		
	1000 pF 5 Hz to 1 kHz 1 kHz to 1 MHz 1 MHz to 13 MHz	0.0028 0.0002 0.0011		
	10 nF 1 kHz	0.0002		
	100 nF 1 kHz	0.0003		
	1 μF 1 kHz	0.0005		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
FREQUENCY				United Kingdom
Specific Values	1 MHz, 5 MHz and 10 MHz	1×10^{-12}		
Other Values	0.1 Hz to 500 MHz 500 MHz to 1 GHz 1 GHz to 2 GHz 2 GHz to 40 GHz	6 in 10^{11} + 0.02 nHz 2 in 10^9 1 in 10^9 5 in 10^{10}		
TIME INTERVAL	10 ns to 10 s	35 ps to 0.12 ns		
PULSE CHARACTERISTICS				
Amplitude	-500 mV to +500 mV	0.1% + 2 mV	Higher amplitude pulses (up to 10 V) can be measured with increased uncertainty.	
Time interval (including rise and fall time)	20 ps to 1 μ s	0.1% + 10 ps		
AMPLITUDE MODULATION				
For the calibration of signal sources and modulation meters between 10% and 90% modulation depth.	Carrier: 0.15 MHz to 10 MHz Modulation: 20 Hz to 400 Hz 400 Hz to 10 kHz 10 kHz to 100 kHz	(% of reading) 1.5% to 3% 1.3% to 2.9% 2.1% to 3.4%	The uncertainties are with respect to the modulation depth and are appropriate for signals with harmonic distortion not exceeding -40 dBc. The uncertainties may be increased for higher levels of distortion at modulation depths less than 30%.	
	Carrier: 10 MHz to 26.5 GHz Modulation: 20 Hz to 400 Hz 400 Hz to 10 kHz 10 kHz to 100 kHz	(% of reading) 1.1% to 2.8% 0.8% to 2.7% 1.8% to 3.1%		
FREQUENCY MODULATION				
For the calibration of signal sources and modulation meters at frequency deviations between 1 kHz and 320 kHz.	Carrier: 10 MHz to 26.5 GHz Modulation: 20 Hz to 100 kHz	2%		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code	
DISTORTION	<p><i>From 20 Hz to 20 kHz:</i> < 0.05% 0.05% to 10%</p> <p><i>From 20 Hz to 50 kHz:</i> < 0.1% 0.1% to 10%</p> <p><i>From 50 kHz to 100 kHz:</i> < 0.18% 0.18% to 10%</p>	<p>0.007% absolute 13.5% of reading</p> <p>0.027% absolute 26.8% of reading</p> <p>0.05% absolute 26.8% of reading</p>		United Kingdom	
RF POWER					
Reference sources					
Output power	1 mW 50 MHz	0.22%	For sources with female type-N connectors		
Source reflection coefficient	0 to 0.1	0.012	Where the output reflection phase is known to be $0^\circ \pm 40^\circ$ or $180^\circ \pm 40^\circ$.		
Waveguide 22 sources	0.2 μ W to 100 mW 26.5 GHz to 40 GHz	2% to 4.7%			
For the calibration of sources and receivers.	2 MHz to 26.5 GHz: 0 dBm to -50 dBm -50 dBm to -90 dBm -90 dBm to -127 dBm	0.15 dB 0.25 dB 0.35 dB	For EMC receivers the results may also be presented in terms of dB μ V in a 50 Ω system.		
CALIBRATION OF RF POWER SOURCES					
The following <i>Best Measurement Capability</i> is for the calibration 75 Ω RF power sources fitted with Type N coaxial connectors. The uncertainties are for the measurement of perfectly matched sources with connectors that are in good condition. Instruments that are not well matched, fitted with different connectors to those shown, or which have connectors that are in poor condition will be assigned larger uncertainties.					
75 Ω Type N coaxial system	1 μ W to 100 μ W	100 μ W to 10 mW	10 mW to 100 mW		Using Model 8483A sensor
100 kHz to 300 kHz	3.1% + 92 nW	3.1%	3.1% + 0.0004 P^2	P = measured power in mW	
300 kHz to 2 GHz	2.5% + 92 nW	2.5%	2.5% + 0.0004 P^2	P = measured power in mW	



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
<p>The following <i>Best Measurement Capability</i> is for the calibration of 50 Ω RF power sources fitted with Type N or APC-3.5 coaxial connectors. The uncertainties are for the measurement of perfectly matched sources with connectors that are in good condition. Instruments that are not well matched, fitted with different connectors to those shown, or which have connectors that are in poor condition will be assigned larger uncertainties.</p>				
50 Ω Type N coaxial system	1 μ W to 100 μ W	100 μ W to 10 mW	10 mW to 100 mW	Using Model 8481A sensor
10 MHz to 2 GHz 2 GHz to 12 GHz 12 GHz to 18 GHz	2.0% + 92 nW 2.0% + 92 nW 3.0% + 92 nW	2.0% 2.6% 3.0%	2.0% + 0.0004 P^2 2.6% + 0.0004 P^2 3.0% + 0.0004 P^2	P = measured power in mW P = measured power in mW P = measured power in mW
50 Ω Type N coaxial system	1 μ W to 100 μ W	100 μ W to 10 mW	10 mW to 100 mW	Using Model 8482A sensor
100 kHz to 4.2 GHz	2.2% + 92 nW	2.2%	2.0% + 0.0004 P^2	P = measured power in mW
50 Ω Type N coaxial system	100 pW to 10 nW	10 nW to 1 μ W	1 μ W to 10 μ W	Using Model 8481D sensor
10 MHz to 4 GHz 4 GHz to 18 GHz	3.3% + 9.2 pW 3.8% + 9.2 pW	3.3% 3.8%	3.3% + 0.0014 $P^{1.9}$ 3.8% + 0.0014 $P^{1.9}$	P = measured power in μ W P = measured power in μ W
50 Ω Type N coaxial system	100 μ W to 10 mW	10 mW to 700 mW	700 mW to 3 W	Using Model 8481H sensor
10 MHz to 2 GHz 2 GHz to 12 GHz 12 GHz to 18 GHz	2.7% + 9.2 μ W 3.1% + 9.2 μ W 3.5% + 9.2 μ W	2.7% 3.1% 3.5%	2.7% + 0.04 $P^{1.4}$ 3.1% + 0.04 $P^{1.4}$ 3.5% + 0.04 $P^{1.4}$	P = measured power in W P = measured power in W P = measured power in W
50 Ω Type N coaxial system	100 μ W to 10 mW	10 mW to 700 mW	700 mW to 3 W	Using Model 8482H sensor
100 kHz to 4.2 GHz	2.8% + 9.2 μ W	2.8%	2.8% + 0.04 $P^{1.4}$	P = measured power in W
50 Ω APC-3.5 coaxial system	1 μ W to 100 μ W	100 μ W to 10 mW	10 mW to 100 mW	Using Model 8485A sensor
50 MHz to 4 GHz 4 GHz to 12 GHz 12 GHz to 18 GHz 18 GHz to 26.5 GHz	2.0% + 92 nW 2.7% + 92 nW 3.4% + 92 nW 3.8% + 92 nW	2.0% 2.7% 3.4% 3.8%	2.0% + 0.0004 P^2 2.7% + 0.0004 P^2 3.4% + 0.0004 P^2 3.8% + 0.0004 P^2	P = measured power in mW P = measured power in mW P = measured power in mW P = measured power in mW
50 Ω APC-3.5 coaxial system	100 pW to 10 nW	10 nW to 1 μ W	1 μ W to 10 μ W	Using Model 8485D sensor
50 MHz to 4 GHz 4 GHz to 12 GHz 12 GHz to 18 GHz 18 GHz to 26.5 GHz	3.3% + 9.2 pW 3.6% + 9.2 pW 4.1% + 9.2 pW 4.5% + 9.2 pW	3.3% 3.6% 4.1% 4.5%	3.3% + 0.0014 $P^{1.9}$ 3.6% + 0.0014 $P^{1.9}$ 4.1% + 0.0014 $P^{1.9}$ 4.5% + 0.0014 $P^{1.9}$	P = measured power in μ W P = measured power in μ W P = measured power in μ W P = measured power in μ W

United Kingdom



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code			
CALIBRATION FACTOR RF Power Sensors							
<p>The table overleaf shows the <i>Best Measurement Capability</i> in % for the various RF power sensors listed below. These are examples and are not intended to be restrictive; other sensors of similar frequency range, power range etc. may also be calibrated.</p> <p>The uncertainties are for devices with values of voltage reflection coefficient (VRC) not exceeding 0.01. The uncertainties will be increased for devices with significantly greater VRC.</p> <p>The uncertainties apply to a measurement procedure that includes an evaluation of the connector repeatability of the device being calibrated. An alternative measurement procedure is available based on a single measurement of the device and generic data for the connector repeatability. Where this procedure is used, the reported uncertainty will be larger.</p>							
Column	Sensor type	Impedance	Connector	Maximum power	Frequency range	Example type	United Kingdom
A	Thermocouple	50 Ω	Type N	100 mW	10 MHz to 18 GHz	8481A	
B	Thermocouple	50 Ω	Type N	3 W	10 MHz to 18 GHz	8481H	
C	Thermocouple	50 Ω	Type N	100 mW	100 kHz to 4.2 GHz	8482A	
D	Thermocouple	50 Ω	Type N	3 W	100 kHz to 4.2 GHz	8482H	
E	Thermocouple	75 Ω	Type N	100 mW	100 kHz to 3 GHz	8483A	
F	Diode	50 Ω	Type N	10 μ W	10 MHz to 18 GHz	8481D	
G	Thermocouple	50 Ω	3.5 mm	100 mW	10 MHz to 26.5 GHz	8485A	
H	Diode	50 Ω	3.5 mm	10 μ W	10 MHz to 26.5 GHz	8485D	
I	Thermistor Mount	50 Ω	Type N	10 mW	10 MHz to 18 GHz	8478B	
J	Thermistor Mount	50 Ω	Type N	10 mW	1 MHz to 1 GHz	478A-H55	
K	Thermocouple	50 Ω	Type N	100 mW	100 kHz to 2.6 GHz	11722A	
L	Thermocouple	50 Ω	Type N	100 mW	10 MHz to 18 GHz	11792A opt. 001	
M	Thermocouple	50 Ω	Type N	100 mW	10 MHz to 26.5 GHz	11792A	



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Calibration Factor - Best Measurement Capability in % for the power sensor types described on Page 13															
Frequency	A	B	C	D	E	F	G	H	I	J	K	L	M	United Kingdom	
100 kHz			0.62	2.1	2.5						1.9				
300 kHz			0.43	1.9	1.3						1.6				
500 kHz			0.41	1.8							1.6				
1 MHz			0.39	1.8	1.0					0.6	1.6				
3 MHz			0.39	1.8	1.0					0.6	1.5				
5 MHz			0.39	1.8							1.5				
10 MHz	0.39	1.9	0.39	1.8	1.0	2.2	0.80	2.3	1.2	0.6	1.5	1.7	2.8		
30 MHz	0.39	1.8	0.39	1.8	1.0	2.1	0.48	2.1	0.6	0.6	1.5	1.6	1.7		
50 MHz									0.6	0.6					
100 MHz	0.40	1.8	0.40	1.8	1.0	2.1	0.47	2.1	0.6	0.6	1.6	1.6	1.6		
200 MHz	0.48		0.48												
300 MHz	0.48	1.8	0.48	1.8	1.0	2.1	0.48	2.1	0.6	0.6	1.6	1.6	1.6		
500 MHz	0.48	1.8	0.48	1.8	1.0	2.1			0.6	0.6	1.6	1.6			
650 MHz	0.48		0.48												
800 MHz	0.48		0.48												
1 GHz	0.47	1.8	0.47	1.8	1.3	2.1	0.48	2.1	0.7	0.9	1.6	1.6	1.6		
1.5 GHz	0.59	1.9	0.59	1.9	1.6	2.2	0.65	2.1	0.7		1.6	1.6	1.6		
1.8 GHz	0.61		0.61												
2 GHz	0.59	1.9	0.59	1.9	1.6	2.2	0.65	2.1	0.9		1.7	1.7	1.6		
2.5 GHz	0.61		0.61		1.7										
2.6 GHz	0.62		0.62	2.0							1.7				
3 GHz	0.64	2.0	0.64	2.0	1.9	2.2	0.65	2.2	1.0			1.7	1.8		
3.5 GHz	0.71		0.71												
3.7 GHz	0.74		0.74												
4 GHz	0.48	2.0	0.48	2.0		2.3	0.66	2.3	1.1			1.7	1.8		
4.2 GHz	0.69		0.69	2.0											
5 GHz	0.57	2.0				2.3	0.66	2.3	1.2			1.8	1.8		
6 GHz	0.57	2.2				2.4	0.84	2.4	1.5			2.0	2.0		
7 GHz	0.57	2.2				2.4	0.84	2.4	1.5			2.0	2.0		
8 GHz	0.57	2.3				2.5	0.94	2.4	1.4			2.0	2.0		
9 GHz	0.67	2.1				2.4	1.2	2.4	1.2			1.9	2.0		
10 GHz	0.67	2.2				2.4	1.2	2.4	1.3			1.9	2.0		
11 GHz	0.68	2.2				2.4	1.2	2.5	1.3			1.9	2.1		
12 GHz	0.66	2.1				2.4	1.2	2.4	1.2			1.9	2.0		
13 GHz	0.66	2.2				2.5	1.2	2.5	1.4			2.0	2.2		
14 GHz	0.66	2.2				2.5	1.2	2.5	1.4			2.0	2.2		
15 GHz	0.76	2.2				2.5	1.2	2.5	1.5			2.0	2.3		
16 GHz	0.80	2.2				2.5	1.3	2.5	1.5			2.0	2.3		
17 GHz	0.93	2.3				2.5	1.3	2.6	1.6			2.1	2.3		
18 GHz	1.04	2.3				2.5	1.3	2.9	1.8			2.1	2.6		
18.5 GHz							1.8	2.9					2.6		
19 GHz							1.6	2.9					2.6		
19.5 GHz							1.7	2.9					2.6		
20 GHz							1.7	2.9					2.6		
20.5 GHz							1.8	2.9					2.5		
21 GHz							1.8	2.9					2.5		
21.5 GHz							1.9	2.9					2.5		
22 GHz							1.9	2.9					2.5		
22.5 GHz							2.0	2.9					2.5		
23 GHz							1.8	2.9					2.5		
23.5 GHz							1.8	2.9					2.5		
24 GHz							1.7	2.9					2.6		
24.5 GHz							1.7	2.9					2.6		
25 GHz							1.6	2.9					2.6		
25.5 GHz							1.7	2.9					2.6		
26 GHz							1.6	2.9					2.7		
26.5 GHz							1.6	3.0					2.7		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code																																																																																																																																		
RF VOLTAGE Specific Values				United Kingdom																																																																																																																																		
<p>The table below shows the <i>Best Measurement Capability</i> in % for the calibration of thermal voltage converters fitted with BNC male connectors. The capability is for 50 Ω, 75 Ω and 600 Ω converters, subject to the following frequency constraints:</p> <p>0.5 V, 75 Ω converters can be calibrated up to a maximum frequency of 30 MHz. 3 V, 600 Ω converters can be calibrated up to a maximum frequency of 10 MHz.</p> <p>Intermediate frequencies may be calibrated but the uncertainty will be at least equal to the larger uncertainty of the two adjacent frequency points</p> <table border="1"> <thead> <tr> <th rowspan="3">Frequency</th> <th colspan="5">Nominal voltage of thermal converter</th> </tr> <tr> <th>0.5 V</th> <th>1 V</th> <th>3 V</th> <th>5 V</th> <th>10 V</th> </tr> <tr> <th>50 Ω or 75 Ω</th> <th>50 Ω or 75 Ω</th> <th>50 Ω or 600 Ω</th> <th>600 Ω</th> <th></th> </tr> </thead> <tbody> <tr><td>100 Hz</td><td>0.11</td><td>0.11</td><td>0.11</td><td></td><td></td></tr> <tr><td>10 kHz</td><td>0.11</td><td>0.11</td><td>0.11</td><td></td><td></td></tr> <tr><td>100 kHz</td><td>0.11</td><td>0.11</td><td>0.11</td><td></td><td>0.12</td></tr> <tr><td>500 kHz</td><td>0.12</td><td>0.11</td><td>0.11</td><td></td><td>0.12</td></tr> <tr><td>700 kHz</td><td></td><td></td><td></td><td></td><td>0.12</td></tr> <tr><td>1 MHz</td><td>0.12</td><td>0.12</td><td>0.12</td><td>0.07</td><td>0.12</td></tr> <tr><td>3 MHz</td><td></td><td></td><td></td><td>0.08</td><td>0.12</td></tr> <tr><td>5 MHz</td><td>0.22</td><td>0.22</td><td>0.22</td><td>0.17</td><td>0.22</td></tr> <tr><td>7 MHz</td><td>0.22</td><td>0.22</td><td>0.22</td><td>0.17</td><td>0.22</td></tr> <tr><td>10 MHz</td><td>0.23</td><td>0.22</td><td>0.22</td><td>0.17</td><td>0.22</td></tr> <tr><td>20 MHz</td><td>0.24</td><td>0.23</td><td>0.23</td><td></td><td></td></tr> <tr><td>30 MHz</td><td>0.25</td><td>0.24</td><td>0.24</td><td></td><td></td></tr> <tr><td>40 MHz</td><td>0.36</td><td>0.35</td><td>0.35</td><td></td><td></td></tr> <tr><td>50 MHz</td><td>0.38</td><td>0.36</td><td>0.36</td><td></td><td></td></tr> <tr><td>60 MHz</td><td>0.40</td><td>0.37</td><td>0.37</td><td></td><td></td></tr> <tr><td>70 MHz</td><td>0.60</td><td>0.57</td><td>0.57</td><td></td><td></td></tr> <tr><td>80 MHz</td><td>0.62</td><td>0.58</td><td>0.58</td><td></td><td></td></tr> <tr><td>90 MHz</td><td>0.64</td><td>0.59</td><td>0.59</td><td></td><td></td></tr> <tr><td>100 MHz</td><td>0.66</td><td>0.60</td><td>0.60</td><td></td><td></td></tr> </tbody> </table>					Frequency	Nominal voltage of thermal converter					0.5 V	1 V	3 V	5 V	10 V	50 Ω or 75 Ω	50 Ω or 75 Ω	50 Ω or 600 Ω	600 Ω		100 Hz	0.11	0.11	0.11			10 kHz	0.11	0.11	0.11			100 kHz	0.11	0.11	0.11		0.12	500 kHz	0.12	0.11	0.11		0.12	700 kHz					0.12	1 MHz	0.12	0.12	0.12	0.07	0.12	3 MHz				0.08	0.12	5 MHz	0.22	0.22	0.22	0.17	0.22	7 MHz	0.22	0.22	0.22	0.17	0.22	10 MHz	0.23	0.22	0.22	0.17	0.22	20 MHz	0.24	0.23	0.23			30 MHz	0.25	0.24	0.24			40 MHz	0.36	0.35	0.35			50 MHz	0.38	0.36	0.36			60 MHz	0.40	0.37	0.37			70 MHz	0.60	0.57	0.57			80 MHz	0.62	0.58	0.58			90 MHz	0.64	0.59	0.59			100 MHz	0.66	0.60	0.60		
Frequency	Nominal voltage of thermal converter																																																																																																																																					
	0.5 V	1 V	3 V			5 V	10 V																																																																																																																															
	50 Ω or 75 Ω	50 Ω or 75 Ω	50 Ω or 600 Ω		600 Ω																																																																																																																																	
100 Hz	0.11	0.11	0.11																																																																																																																																			
10 kHz	0.11	0.11	0.11																																																																																																																																			
100 kHz	0.11	0.11	0.11			0.12																																																																																																																																
500 kHz	0.12	0.11	0.11			0.12																																																																																																																																
700 kHz						0.12																																																																																																																																
1 MHz	0.12	0.12	0.12		0.07	0.12																																																																																																																																
3 MHz					0.08	0.12																																																																																																																																
5 MHz	0.22	0.22	0.22		0.17	0.22																																																																																																																																
7 MHz	0.22	0.22	0.22		0.17	0.22																																																																																																																																
10 MHz	0.23	0.22	0.22		0.17	0.22																																																																																																																																
20 MHz	0.24	0.23	0.23																																																																																																																																			
30 MHz	0.25	0.24	0.24																																																																																																																																			
40 MHz	0.36	0.35	0.35																																																																																																																																			
50 MHz	0.38	0.36	0.36																																																																																																																																			
60 MHz	0.40	0.37	0.37																																																																																																																																			
70 MHz	0.60	0.57	0.57																																																																																																																																			
80 MHz	0.62	0.58	0.58																																																																																																																																			
90 MHz	0.64	0.59	0.59																																																																																																																																			
100 MHz	0.66	0.60	0.60																																																																																																																																			
Other Values	0.25 V to 5 V (10MHz max. for 3V to 5V) 20 Hz to 1 MHz 1 MHz to 10 MHz 10 MHz to 30 MHz 30 MHz to 60 MHz 60 MHz to 80 MHz 80 MHz to 100 MHz 5 V to 10 V 20 Hz to 3 MHz 3 MHz to 10 MHz	Generation 0.15% 0.31% 0.31% 0.50% 0.82% 0.82% 0.16% 0.26%	Measurement 0.32% 0.39% 0.59% 0.70% 0.89% 1.3% 0.16% 0.26%	For calibration of RF voltmeters and measurement of RF sources, using thermal voltage converters.																																																																																																																																		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RF ATTENUATION Specific Values For the calibration of spectrum analysers, receivers, network analysers, signal generators etc.	<p><i>60 MHz:</i> 10, 20, 30 dB 40, 50, 60, 70 dB 80, 90, 100, 110 dB</p> <p><i>3 MHz:</i> 2, 4, 6, 8 and 10 dB 20 and 30 dB 40 dB 50 dB 60 dB 70 and 80 dB</p> <p><i>10 MHz:</i> 38 dB 58 dB 98 dB</p> <p><i>40 MHz:</i> 38 dB 58 dB 98 dB</p> <p><i>80 MHz:</i> 30 dB 58 dB 98 dB</p>	<p>0.005 dB 0.006 dB 0.006 dB</p> <p>0.013 dB 0.014 dB 0.015 dB 0.016 dB 0.017 dB 0.018 dB</p> <p>0.015 dB 0.016 dB 0.023 dB</p> <p>0.015 dB 0.017 dB 0.023 dB</p> <p>0.016 dB 0.017 dB 0.023 dB</p>	<p>All values are relative to a nominal level of 0 dBm.</p> <p>The uncertainties apply at the stated frequencies. Measurements can also be made at intermediate frequencies within the range, with increased uncertainty.</p>	United Kingdom
Other Values (50 Ω coaxial system)	<p><i>300 kHz to 6 GHz:</i> 0 dB to 40 dB 40 dB to 50 dB 50 dB to 60 dB 60 dB to 70 dB 70 dB to 75 dB</p> <p>0 dB to 90 dB <i>50 MHz to 26.5 GHz</i></p>	<p>0.03 dB to 0.08 dB 0.04 dB to 0.13 dB 0.08 dB to 0.32 dB 0.20 dB to 0.95 dB 0.93 dB to 1.7 dB</p> <p>See table on following page</p>	<p>The uncertainties stated are for two-port 50 Ω devices fitted with APC7, Type-N or APC3.5 connectors that have input and output VRC of less than 0.1. The uncertainties may be increased for higher values of VRC.</p>	



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code					
<p>The following table gives the Best Measurement Capability in dB for ranges of measured attenuation using a modified version of a Vector Network Analyser Model 8510C. The uncertainties are for two-port 50 Ω devices fitted with APC-7, Type N or APC-3.5 connectors that have input and output VRC of less than 0.1. The uncertainties may be increased for higher values of VRC.</p>									
Frequency GHz	Uncertainty in measured attenuation (dB) for the ranges shown								
	0 dB to 10 dB	10 dB to 20 dB	20 dB to 30 dB	30 dB to 40 dB	40 dB to 50 dB	50 dB to 60 dB	60 dB to 70 dB	70 dB to 80 dB	80 dB to 90 dB
0.05	0.05	0.05	0.05	0.05	0.05	0.06	0.11	0.32	1.0
0.1	0.05	0.05	0.05	0.05	0.05	0.06	0.11	0.33	1.0
0.3	0.05	0.05	0.05	0.05	0.05	0.06	0.11	0.33	1.0
1	0.05	0.05	0.05	0.05	0.05	0.06	0.11	0.33	1.0
1.5	0.05	0.05	0.05	0.05	0.05	0.06	0.12	0.34	1.1
2	0.05	0.05	0.05	0.05	0.06	0.06	0.12	0.36	1.1
3	0.05	0.05	0.05	0.05	0.06	0.07	0.13	0.39	1.2
4	0.05	0.05	0.05	0.05	0.06	0.07	0.14	0.42	1.3
5	0.05	0.05	0.05	0.06	0.06	0.07	0.15	0.45	1.4
6	0.05	0.05	0.05	0.06	0.06	0.08	0.16	0.49	1.5
7	0.05	0.05	0.06	0.06	0.06	0.08	0.18	0.53	1.6
8	0.05	0.06	0.06	0.06	0.06	0.08	0.19	0.58	1.7
9	0.06	0.06	0.06	0.06	0.07	0.09	0.21	0.63	1.9
10	0.06	0.06	0.06	0.06	0.07	0.09	0.23	0.69	2.1
11	0.06	0.06	0.06	0.06	0.07	0.10	0.25	0.75	2.2
12	0.06	0.06	0.06	0.07	0.07	0.10	0.27	0.81	2.4
13	0.06	0.06	0.06	0.07	0.07	0.11	0.29	0.89	2.6
14	0.06	0.06	0.07	0.07	0.08	0.12	0.32	0.97	2.9
15	0.06	0.06	0.07	0.07	0.08	0.13	0.35	1.1	3.1
16	0.06	0.07	0.07	0.07	0.08	0.14	0.38	1.2	3.3
17	0.07	0.07	0.07	0.07	0.09	0.15	0.42	1.3	3.6
18	0.07	0.07	0.07	0.08	0.09	0.16	0.45	1.4	3.9
19	0.07	0.07	0.07	0.08	0.09	0.17	0.49	1.5	4.2
20	0.07	0.07	0.07	0.08	0.10	0.19	0.54	1.6	4.6
21	0.07	0.07	0.08	0.08	0.10	0.20	0.59	1.8	4.9
22	0.07	0.07	0.08	0.08	0.10	0.22	0.64	1.9	5.3
23	0.07	0.08	0.08	0.08	0.11	0.24	0.70	2.1	5.7
24	0.07	0.08	0.08	0.09	0.11	0.26	0.76	2.2	6.1
25	0.08	0.08	0.08	0.09	0.12	0.28	0.82	2.4	6.5
26	0.08	0.08	0.08	0.09	0.13	0.30	0.90	2.6	7.0
26.5	0.08	0.08	0.08	0.09	0.13	0.32	0.94	2.7	7.2
				United Kingdom					



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
DIRECTIVITY of VRC bridges				United Kingdom
The capability described below is for the measurement of directivity on 50 Ω VRC bridges and similar devices. The uncertainties are given in linear quantities (VRC) where the range applies to the range of measured directivity. The values and uncertainties may be reported in terms of return loss (dB), calculated from the linear values and uncertainties.				
BNC connectors	16 dB to 50 dB <i>0.1 MHz to 110 MHz</i>	0.017 to 0.006		
Type N connectors	16 dB to 50 dB <i>0.1 MHz to 2 GHz</i> <i>2 GHz to 18 GHz</i>	0.017 to 0.006 0.019 to 0.010		
3.5 mm connectors	16 dB to 50 dB <i>10 MHz to 3 GHz</i> <i>3 GHz to 18 GHz</i> <i>18 GHz to 26.5 GHz</i>	0.018 to 0.006 0.018 to 0.008 0.019 to 0.008		
7 mm connectors	16 dB to 50 dB <i>10 MHz to 18 GHz</i>	0.017 to 0.004		
2.4 mm connectors	16 dB to 50 dB <i>10 MHz to 4 GHz</i> <i>4 GHz to 20 GHz</i> <i>20 GHz to 36 GHz</i> <i>36 GHz to 40 GHz</i> <i>40 GHz to 50 GHz</i>	0.019 to 0.009 0.019 to 0.010 0.021 to 0.013 0.023 to 0.016 0.026 to 0.020		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code													
CISPR16 Pulse Response	Nominal level	<i>PRF</i>		United Kingdom													
	64 dB μ V	100 Hz	0.55 dB		United Kingdom												
	63 dB μ V	60 Hz	0.55 dB			United Kingdom											
	56 dB μ V	10 Hz	0.55 dB				United Kingdom										
	52.5 dB μ V	5 Hz	0.55 dB					United Kingdom									
	47 dB μ V	2 Hz	0.55 dB						United Kingdom								
	43 dB μ V	1 Hz	0.55 dB							United Kingdom							
	53.5 dB μ V	20 Hz	0.55 dB								United Kingdom						
	50 dB μ V	10 Hz	0.55 dB									United Kingdom					
	39.5 dB μ V	2 Hz	0.55 dB										United Kingdom				
	37.5 dB μ V	1 Hz	0.55 dB											United Kingdom			
	51 dB μ V	20 Hz	0.55 dB												United Kingdom		
	46 dB μ V	10 Hz	0.55 dB													United Kingdom	
	34 dB μ V	2 Hz	0.55 dB														United Kingdom
	31.5 dB μ V	1 Hz	0.70 dB														
VOLTAGE REFLECTION COEFFICIENT	50 Ω coaxial system	100 kHz to 300 kHz:		United Kingdom													
		0 to 0.1	0.03 to 0.04		United Kingdom												
		0.1 to 1.0	0.04 to 0.11			United Kingdom											
		300 kHz to 6 GHz:					United Kingdom										
		0 to 0.2	0.005 to 0.01					United Kingdom									
		0.2 to 0.4	0.007 to 0.015						United Kingdom								
		0.4 to 0.6	0.011 to 0.023							United Kingdom							
		0.6 to 0.8	0.018 to 0.033								United Kingdom						
		0.8 to 1.0	0.027 to 0.047									United Kingdom					
		50 MHz to 26.5 GHz:											United Kingdom				
0 to 1.0	See table on following page	United Kingdom															



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code

Frequency GHz	Uncertainty in Voltage Reflection Coefficient Magnitude for the ranges shown									
	0 to 0.1	0.1 to 0.2	0.2 to 0.3	0.3 to 0.4	0.4 to 0.5	0.5 to 0.6	0.6 to 0.7	0.7 to 0.8	0.8 to 0.9	0.9 to 1.0
0.05	0.007	0.007	0.007	0.007	0.008	0.009	0.010	0.012	0.016	0.021
0.1	0.007	0.007	0.007	0.007	0.008	0.009	0.010	0.012	0.016	0.021
0.3	0.007	0.007	0.007	0.007	0.008	0.009	0.010	0.012	0.016	0.021
1	0.007	0.007	0.007	0.008	0.008	0.009	0.010	0.012	0.016	0.022
1.5	0.008	0.008	0.008	0.008	0.008	0.009	0.010	0.013	0.016	0.022
2	0.008	0.008	0.008	0.008	0.009	0.009	0.011	0.013	0.017	0.023
3	0.008	0.008	0.008	0.009	0.009	0.010	0.011	0.014	0.018	0.023
4	0.009	0.009	0.009	0.009	0.009	0.010	0.012	0.014	0.018	0.024
5	0.009	0.009	0.009	0.009	0.010	0.011	0.012	0.015	0.019	0.025
6	0.010	0.010	0.010	0.010	0.010	0.011	0.013	0.015	0.019	0.026
7	0.010	0.010	0.010	0.010	0.011	0.012	0.013	0.016	0.020	0.026
8	0.010	0.010	0.010	0.011	0.011	0.012	0.014	0.016	0.021	0.027
9	0.011	0.011	0.011	0.011	0.011	0.012	0.014	0.017	0.021	0.028
10	0.011	0.011	0.011	0.011	0.012	0.013	0.015	0.017	0.022	0.028
11	0.012	0.012	0.012	0.012	0.012	0.013	0.015	0.018	0.022	0.029
12	0.012	0.012	0.012	0.012	0.013	0.014	0.015	0.018	0.023	0.030
13	0.012	0.012	0.012	0.013	0.013	0.014	0.016	0.019	0.024	0.031
14	0.013	0.013	0.013	0.013	0.014	0.015	0.016	0.019	0.024	0.031
15	0.013	0.013	0.013	0.013	0.014	0.015	0.017	0.020	0.025	0.032
16	0.014	0.014	0.014	0.014	0.014	0.015	0.017	0.020	0.025	0.033
17	0.014	0.014	0.014	0.014	0.015	0.016	0.018	0.021	0.026	0.033
18	0.014	0.014	0.014	0.015	0.015	0.016	0.018	0.022	0.027	0.034
19	0.014	0.014	0.014	0.015	0.015	0.016	0.018	0.022	0.027	0.035
20	0.014	0.014	0.015	0.015	0.015	0.016	0.018	0.022	0.027	0.035
21	0.014	0.014	0.015	0.015	0.015	0.016	0.018	0.022	0.027	0.035
22	0.014	0.014	0.015	0.015	0.015	0.016	0.019	0.022	0.027	0.036
23	0.014	0.014	0.015	0.015	0.015	0.017	0.019	0.022	0.028	0.036
24	0.015	0.015	0.015	0.015	0.015	0.017	0.019	0.022	0.028	0.036
25	0.015	0.015	0.015	0.015	0.015	0.017	0.019	0.022	0.028	0.037
26	0.015	0.015	0.015	0.015	0.015	0.017	0.019	0.023	0.028	0.037
26.5	0.015	0.015	0.015	0.015	0.015	0.017	0.019	0.023	0.028	0.037

United Kingdom

VSWR 75 Ω coaxial system	1.0 to 1.2 <i>100 kHz to 500 kHz</i> <i>500 kHz to 10 MHz</i> <i>10 MHz to 40 MHz</i> <i>40 MHz to 2 GHz</i>	0.06 to 0.17 0.06 to 0.10 0.07 to 0.12 0.07 to 0.11	For devices fitted with 75 Ω Type N connectors.
	1.2 to 2.0 <i>100 kHz to 500 kHz</i> <i>500 kHz to 10 MHz</i> <i>10 MHz to 40 MHz</i> <i>40 MHz to 2 GHz</i>	0.17 to 0.47 0.10 to 0.32 0.12 to 0.36 0.11 to 0.33	For devices fitted with 75 Ω Type N connectors.




0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code 
<p align="center">Internet Primary Impedance Measurement System (#PIMMS)</p> <p>#PIMMS is an Internet-enabled measurement service supplied and maintained by the UK's National Physical Laboratory. Measurements are performed using a Microwave Network Analyser connected via the Internet to the NPL #PIMMS server.</p> <p>NOTES</p> <p>For the linear voltage reflection and transmission coefficient measurands (i.e. complex-valued S-parameters) described in this section of the schedule, the Best Measurement Capability is shown as an interval of values, where a selected value within the interval represents an expanded uncertainty at a level of confidence of approximately 95%. Furthermore, a selected value within the interval will represent the uncertainty applied equally and simultaneously to <i>both</i> the Real and Imaginary parts of the S-parameter. The uncertainty value therefore defines a circular region of uncertainty, in the appropriate S-parameter plane, centred on the measured, quoted, mean value with a radius equal to the stated expanded uncertainty.</p> <p>For Voltage Reflection Coefficients (VRCs), the stated uncertainty is assumed here to be independent of the nominal VRC , so a single interval is presented applicable for all VRC in the range $0 \leq VRC \leq 1$. For Voltage Transmission Coefficients (VTCs), the stated uncertainty is dependent on the nominal VTC , so uncertainty intervals are presented for selected, representative, values of VTC in the range $0 \leq VTC \leq 1$.</p> <p>The measured values and uncertainties may also be reported as dB, VSWR, or VRC, with the angle reported in degrees or radians. Typical examples are presented in this schedule for clarification purposes.</p>				United Kingdom
<p align="center">#PIMMS</p> <p>LINEAR COMPLEX VOLTAGE REFLECTION COEFFICIENT (VRC) IN 50 Ω COAXIAL LINE</p>	<p>45 MHz to 18 GHz: $-1 \leq \text{Re}(\text{VRC}) \leq +1$ $-1 \leq \text{Im}(\text{VRC}) \leq +1$ constrained by: $0 \leq \text{VRC} \leq 1$</p> <p>45 MHz to 26.5 GHz: $-1 \leq \text{Re}(\text{VRC}) \leq +1$ $-1 \leq \text{Im}(\text{VRC}) \leq +1$ constrained by: $0 \leq \text{VRC} \leq 1$</p>	<p>0.002 to 0.007</p> <p>0.006 to 0.012</p>	<p>For devices fitted with 7mm or Type-N connectors only.</p> <p>For devices fitted with 3.5mm connectors.</p>	
<p align="center">#PIMMS</p> <p>LINEAR COMPLEX VOLTAGE TRANSMISSION COEFFICIENT (VTC) IN 50 Ω COAXIAL LINE</p>	<p>45 MHz to 18 GHz: $-1 \leq \text{Re}(\text{VTC}) \leq +1$ $-1 \leq \text{Im}(\text{VTC}) \leq +1$ constrained by: $0 \leq \text{VTC} \leq 1$</p> <p> VTC = 0 VTC = 0.1 VTC = 1</p>	<p>0.000005 to 0.00005 0.00015 to 0.0005 0.0015 to 0.005</p>	<p>For devices fitted with 7mm or Type-N connectors only.</p>	



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
<p align="center">PIMMS</p> <p>LINEAR COMPLEX VOLTAGE TRANSMISSION COEFFICIENT (VTC) IN 50 Ω COAXIAL LINE</p> <p>Typical examples expressed in terms of attenuation and transmission phase</p>	<p>Attenuation from 0 dB to 40 dB</p> <p>Attenuation at 50 dB</p> <p>Attenuation at 60 dB</p> <p>Attenuation at 70 dB</p> <p>Attenuation at 80 dB</p> <p>45 MHz to 26.5 GHz: $-1 \leq \text{Re}(VTC) \leq +1$ $-1 \leq \text{Im}(VTC) \leq +1$ constrained by: $0 \leq VTC \leq 1$</p>	<p>0.013 dB to 0.043 dB 0.09° to 0.28°</p> <p>0.014 dB to 0.14 dB 0.09° to 0.91°</p> <p>0.043 dB to 0.42 dB 0.29° to 2.9°</p> <p>0.14 dB to 1.3 dB 0.91° to 9.1°</p> <p>0.42 dB to 3.5 dB 2.9° to 30°</p>	<p>For devices fitted with 7mm or Type-N connectors only.</p> <p>For devices fitted with 3.5mm connectors.</p>	United Kingdom
<p>Typical examples expressed in terms of attenuation and transmission phase</p>	<p>Attenuation from 0 dB to 30 dB</p> <p>Attenuation at 40 dB</p> <p>Attenuation at 50 dB</p> <p>Attenuation at 60 dB</p> <p>Attenuation at 70 dB</p> <p>Attenuation at 80 dB</p>	<p>0.013 dB to 0.043 dB 0.09° to 0.29°</p> <p>0.014 dB to 0.086 dB 0.09° to 0.57°</p> <p>0.014 dB to 0.27 dB 0.09° to 1.8°</p> <p>0.043 dB to 0.83 dB 0.29° to 5.7°</p> <p>0.14 dB to 2.4 dB 0.91° to 18°</p> <p>0.42 dB to 6.0 dB 2.9° to 90°</p>	<p>For devices fitted with 3.5mm connectors</p>	



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
DC RESISTANCE				Site calibration
Measurement	1 Ω to 110 Ω 110 Ω to 110 k Ω 110 k Ω to 1.1 M Ω 1.1 M Ω to 11 M Ω 11 M Ω to 110 M Ω	90 ppm to 30 ppm 30 ppm 60 ppm to 30 ppm 810 ppm to 185 ppm 0.15% to 0.09%		
Generation	0 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	1 m Ω 30 ppm 15 ppm 15 ppm 15 ppm 15 ppm 35 ppm 55 ppm 130 ppm		
DC VOLTAGE				
Measurement	0 to 1000 V	15 ppm + 2 μ V		
Generation	0 to 1000 V	10 ppm + 2 μ V		
DC CURRENT				
Measurement	10 μ A to 110 μ A 110 μ A to 11 mA 11 mA to 110 mA 110 mA to 1.05 A	230 ppm to 80 ppm 170 ppm to 70 ppm 230 ppm to 130 ppm 380 ppm to 210 ppm		
Generation	0.1 μ A to 200 μ A 0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	90 ppm + 2 nA 40 ppm + 10 nA 40 ppm + 100 nA 40 ppm + 1 μ A 90 ppm + 2 μ A		
AC VOLTAGE				
Measurement	1 mV to 11 mV 20 Hz to 40 Hz 40 Hz to 50 kHz 50 kHz to 100 kHz	0.7% 0.3% 0.7%		
	11 mV to 11 V 20 Hz to 50 kHz 50 kHz to 100 kHz	0.05% to 0.02% 0.1%		
	11 V to 110 V 20 Hz to 50 kHz 50 kHz to 100 kHz	0.06% to 0.03% 0.14%		
	110 V to 700 V 20 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.08% to 0.05% 0.14% 0.32%		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AC VOLTAGE (cont'd)				Site calibration
Generation	0.1 mV to 2 mV 20 Hz to 100 kHz	0.25% + 5 μ V		
	2 mV to 200 mV 20 Hz to 100 kHz	0.1% + 5 μ V		
	200 mV to 20 V 20 Hz to 100 kHz 100 kHz to 300 kHz 300 kHz to 1 MHz	0.07% to 0.02% 0.05% to 0.3% 0.5% to 3.5%		
	20 V to 200 V 20 Hz to 30 kHz 30 kHz to 100 kHz	0.05% to 0.02% 0.08% to 0.025%		
	200 V to 1000 V 45 Hz to 30 kHz	0.06% to 0.04%		
	200 V to 750 V 30 kHz to 100 kHz	0.25% to 0.15%		
AC CURRENT				
Measurement	10 μ A to 110 mA 45 Hz to 5 kHz	0.26% to 0.5%		
	110 mA to 1.05 A 45 Hz to 5 kHz	0.12%		
Generation	10 μ A to 1A 20 Hz to 5 kHz	0.15% to 0.03%		
FREQUENCY				
Measurement	0.01 Hz to 100 MHz 100 MHz to 26.5 GHz	2 in 10^{10} to 4 in 10^9 1 in 10^{10} + 1 Hz		
Generation	0.01 Hz to 26.5 GHz	1 in 10^{10}		
TIME INTERVAL	10 ns to 10 s	35 ps to 0.12 ns		
RF VOLTAGE				
	0.25 V to 1 V 100 kHz to 15 MHz 15 MHz to 35 MHz 35 MHz to 55 MHz 55 MHz to 100 MHz	0.13% 0.3% 0.5% 1.1%		
	1.5 V to 3 V 100 kHz to 15 MHz 15 MHz to 35 MHz 35 MHz to 55 MHz 55 MHz to 100 MHz	0.13% 0.3% 0.5% 2.1%		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
AMPLITUDE MODULATION	Modulation Depths between 10% and 90% Carrier: 0.15 MHz to 10 MHz Modulation: 20 Hz to 400 Hz 400 Hz to 10 kHz 10 kHz to 100 kHz Carrier: 10 MHz to 26.5 GHz Modulation: 20 Hz to 400 Hz 400 Hz to 10 kHz 10 kHz to 100 kHz	1.5% to 3.0% of reading 1.3% to 2.9% of reading 2.1% to 3.4% of reading 1.1% to 2.8% of reading 0.8% to 2.7% of reading 1.8% to 3.1% of reading	For the calibration of sources and modulation meters. The uncertainty is with respect to the modulation depth for signals of total harmonic distortion less than -40 dBc. The uncertainty may be increased for higher levels of distortion at modulation depths less than 30%.	Site calibration
FREQUENCY MODULATION	Frequency Deviations between 1 kHz and 320 kHz Carrier: 10 MHz to 26.5 GHz Modulation: 20 Hz to 100 kHz	2.0%	For the calibration of sources and modulation meters.	
DISTORTION	20 Hz to 20 kHz: <0.05% 0.05% to 10% 20 kHz to 50 kHz: <0.1% 0.1% to 10% 50 Hz to 100 kHz: <0.18% 0.18% to 10%	0.007% absolute 13.5% of reading 0.027% absolute 26.8% of reading 0.05% absolute 26.8% of reading	For the calibration of modulation oscillators fitted to RF and microwave sources.	
RF POWER				
Measurement				
Reference Sources	1 mW 50 MHz	0.7%	For sources with female Type N connectors.	
50 Ω system	10 nW to 1 W 100 kHz to 26.5 GHz	2% to 7%	Type N or APC -3.5 connectors.	
75 Ω system	0.2 μ W to 100 mW 100 kHz to 2.0 GHz	2% to 5%	Precision Type N connectors.	
Receiver and signal source calibration	2 MHz to 26.5 GHz: 0 dBm to -50 dBm -50 dBm to -90 dBm -90 dBm to -127 dBm	0.15 dB 0.25 dB 0.35 dB		



0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
CALIBRATION FACTOR 50 Ω system	100 kHz	1.5%	The uncertainties are for Power Sensors and Thermistor Mounts fitted with either precision male Type N or APC 3.5 mm connectors.	Site calibration
	300 kHz	0.9%		
	500 kHz	0.8%	Other connector types will be assigned larger uncertainties.	
	1 MHz to 300 MHz	0.6%		
	400 MHz to 1.0 GHz	0.7%		
	1.5 GHz	0.8%		
	2 GHz	0.9%		
	3 GHz	1.0%		
	4 GHz and 5 GHz	1.1%		
	6 GHz	1.2%		
	7 GHz and 8 GHz	1.5%		
	9 GHz	1.6%		
	10 GHz to 12 GHz	1.7%		
	13 GHz	1.8%		
	14 GHz	1.9%		
	15 GHz	2.0%		
	16 GHz and 17 GHz	2.1%		
	18 GHz	2.3%		
	18.5 GHz to 24.5 GHz	2.7%		
	25 GHz	2.8%		
	25.5 GHz	3.1%		
	26 GHz and 26.5 GHz	3.2%		
75 Ω system	100 kHz	2.9%	The uncertainties are for Power Sensors fitted with either precision male Type N connectors.	
	300 kHz	1.6%		
	1 MHz to 100 MHz	1.4%	The calibration method is based upon one set of measurements being performed with the uncertainty evaluation using a generic value for Type A uncertainties.	
	125 MHz to 150 MHz	1.5%		
	300 MHz	1.6%		
	500 MHz	1.7%		
	1.0 GHz	2.2%		
	1.5 GHz	2.6%		
	2.0 GHz	3.0%		
VOLTAGE REFLECTION COEFFICIENT				
50 Ω system	<i>100 kHz to 26.5 GHz:</i>		For devices fitted with male APC 3.5 or precision Type N male connectors	
	0.0 to 0.1	0.01 to 0.04		
	0.1 to 0.2	0.02 to 0.05		
	0.2 to 0.3	0.03 to 0.06		
	0.3 to 1.0	0.03 to 0.12		
VSWR				
75 Ω system	<i>100 kHz to 2.0 GHz:</i>		For devices fitted with precision Type N male connectors.	
	1.0 to 1.2	0.06 to 0.17		
	1.2 to 1.5	0.12 to 0.27		
	1.5 to 2.0	0.19 to 0.48		




0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
				
DC VOLTAGE				Germany
Specific value	10 V	0.46 ppm	For calibration of 10 V DC references	
Other values	1 μ V to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.7 ppm + 0.26 μ V 1.5 ppm 3.5 ppm 3.5 ppm		
DC CURRENT				
Specific values	1 μ A 10 μ A and 100 μ A 1 mA and 10 mA 100 mA and 1 A 10 A 100 A	19 ppm 14 ppm 9.5 ppm 9.5 ppm 15 ppm 51 ppm		
Other values	0.1 μ A to 1 μ A 1 μ A to 100 μ A 100 μ A to 1 A 1 A to 10 A 10 A to 100 A	45 ppm 43 ppm 12 ppm 43 ppm 65 ppm		
DC RESISTANCE				
	1 m Ω to 100 m Ω 100 m Ω to 100 Ω 100 Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω At 10 M Ω 10 M Ω to 100 M Ω At 100 M Ω 100 M Ω to 1 G Ω 1 G Ω to 10 G Ω	12 ppm 3.0 ppm 3.3 ppm 3.9 ppm 13 ppm 4.8 ppm 18 ppm 13 ppm 130 ppm 480 ppm		




0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
				 Germany
AC VOLTAGE	1 mV to 10 mV 1 Hz to 1 kHz	280 ppm	Using fast DC sampling technique	Germany
	10 mV to 7 V 1 Hz to 100 Hz 100 Hz to 500 Hz 500 Hz to 1 kHz	35 ppm 51 ppm 81 ppm		
	7 V to 700 V 1 Hz to 100 Hz	45 ppm	Using thermal transfer technique	
	0.5 V to 3 V 40 Hz to 100 kHz 100 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	30 ppm 43 ppm 170 ppm 330 ppm		
	5 V to 10 V 40 Hz to 50 kHz 50 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	30 ppm 43 ppm 170 ppm 330 ppm		
	15 V to 30 V 40 Hz to 50 kHz 50 kHz to 200 kHz 200 kHz to 500 kHz 500 kHz to 1 MHz	31 ppm 44 ppm 170 ppm 330 ppm		
	50 V to 100 V 40 Hz to 20 kHz 20 kHz to 100 kHz	28 ppm 32 ppm		
	150 V to 300 V 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	29 ppm 35 ppm 55 ppm		
500 V to 1000 V 40 Hz to 30 kHz	35 ppm			
RF VOLTAGE (Thermal voltage converter calibration)	0.15 V to 3 V 1 MHz to 3 MHz 3 MHz to 30 MHz 30 MHz to 60 MHz 60 MHz to 100 MHz	0.059% 0.10% 0.24% 0.43%	Calibration of TVCs with BNC connectors < 100 MHz or GR874 connectors < 50 MHz The uncertainty may be increased with other connector types.	




0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
				
RF POWER				Germany
Calibration factor				
50 Ω Power Sensors with Type N connectors	- 25 dBm to + 10 dBm	3.0 %		
	100 kHz to 300 kHz	3.0 %	$ \Gamma < 0.15$	
	300 kHz to 10 MHz	2.1 %	$ \Gamma < 0.10$	
	10 MHz to 1 GHz	1.6 %	$ \Gamma < 0.04$	
1 GHz steps:	1 GHz to 5 GHz	2.0 %	$ \Gamma < 0.05$	
1 GHz steps:	5 GHz to 9 GHz	2.6 %	$ \Gamma < 0.06$	
1 GHz steps:	9 GHz to 13 GHz	3.0 %	$ \Gamma < 0.07$	
1 GHz steps:	13 GHz to 16 GHz	3.7 %	$ \Gamma < 0.09$	
1 GHz steps:	16 GHz to 18 GHz	4.7 %	$ \Gamma < 0.09$	
50 Ω Power Sensors with APC3.5 connectors	- 25 dBm to + 10 dBm			
	50 MHz to 1 GHz	1.5 %	$ \Gamma < 0.04$	
1 GHz steps:	1 GHz to 5 GHz	2.0 %	$ \Gamma < 0.05$	
1 GHz steps:	5 GHz to 9 GHz	2.6 %	$ \Gamma < 0.06$	
1 GHz steps:	9 GHz to 13 GHz	3.1 %	$ \Gamma < 0.07$	
1 GHz steps:	13 GHz to 18 GHz	3.7 %	$ \Gamma < 0.09$	
1 GHz steps:	18 GHz to 26.5 GHz	4.6 %	$ \Gamma < 0.10$	
VOLTAGE REFLECTION COEFFICIENT $ \Gamma $				
For devices fitted with 50 Ω Type N connectors	100 kHz to 1 MHz:			
	0.01 to 0.048	0.050		
	0.048 to 0.13	0.060		
	0.13 to 0.2	0.070		
	0.2 to 0.25	0.075		
	1 MHz to 10 MHz:			
	0.01 to 0.048	0.020		
	0.048 to 0.13	0.030		
	0.13 to 0.2	0.040		
	0.2 to 0.25	0.045		
	10 MHz to 8.4 GHz:			
	0.01 to 0.048	0.030		
	0.048 to 0.13	0.035		
	0.13 to 0.2	0.040		
	0.2 to 0.25	0.045		




0147
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Agilent Technologies UK Limited
Issue No: 036 Issue date: 3 October 2008

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code		
						
VOLTAGE REFLECTION COEFFICIENT $ \Gamma $ (continued) For devices fitted with 50 Ω Type N connectors For devices fitted with 50 Ω APC3.5 connectors	<i>8.4 GHz to 12.4 GHz:</i> 0.01 to 0.048 0.048 to 0.13 0.13 to 0.2 0.2 to 0.25	0.030 0.035 0.045 0.050		Germany		
	<i>12 GHz to 18 GHz:</i> 0.01 to 0.048 0.048 to 0.13 0.13 to 0.2 0.2 to 0.25	0.035 0.045 0.055 0.060				
	<i>10 MHz to 8.4 GHz:</i> 0.01 to 0.048 0.048 to 0.13 0.13 to 0.2 0.2 to 0.25	0.020 0.025 0.030 0.040				
	<i>8.4 GHz to 20 GHz:</i> 0.01 to 0.048 0.048 to 0.13 0.13 to 0.2 0.2 to 0.25	0.025 0.030 0.040 0.050				
	<i>20 GHz to 26.5 GHz:</i> 0.01 to 0.048 0.048 to 0.13 0.13 to 0.2 0.2 to 0.25	0.030 0.040 0.055 0.070				
	FREQUENCY					
	Specific values	100 kHz, 1 MHz and 10 MHz	$2.3 \text{ in } 10^{11}$		The <i>Best Measurement Capability</i> is for a measurement period not less than 8 hours.	
	Other values	0.1 mHz to 5 GHz	$2 \cdot \sqrt{(1.2 \cdot 10^{-11})^2 + u_T^2}$		u_T is the trigger uncertainty and will depend on the slew rate and noise levels associated with the signal being measured. It may be particularly significant at low frequencies.	